

CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

DATE OF ISSUE: 14-Jul-14

CERTIFICATE No: 311689



Lambda
CALIBRATION LTD

11-13 Chorley Central
Business Park
Stump Lane
Chorley
PR6 0BL
Tel: 0845 241 1533

Page 1 of 2
APPROVED SIGNATORY

J Carmichael K Quigley
J Carmichael K Quigley
P Davies A Kelly D Whalley

Customer: Premier Autoclaves Service & Solutions Ltd, Keighley BD20 5LN

Item No: 61021066

Description: A used Druck DPI 610 Digital Pressure Calibrator

Date of Cal: 10-Jul-14

Basis: Procedure No. C.I.M-101, C.I.E-01&03 and the Manufacturer's Specifications

Equipment Used: LDP-06, LDW-100-02, LMMC-10

Temperature: 20°C +/- 1.5°C

Calibration Medium: 0-1.1 bar Abs: Air, 1.1-7 bar Abs: Nitrogen

Inlet Port Orientation: Vertical

Pressure Ref Point: The top of the pressure inlet port

Visual Examination: The instrument casing is cracked around the pressure port.
Battery replaced

Notes:

Manufacturers Specifications: Pressure: $\pm 0.025\%$ FSD
Voltage: $\pm(0.05\%$ Rdg + 0.004% FS)
Current: $\pm(0.05\%$ Rdg + 0.004% FS)

Uncertainty of Measurement: 0 to 1.1 bar Abs: ± 0.00012 bar + readability of UUT
1.1 to 7 bar Abs : $\pm 0.0087\%$ of applied pressure + 0.00014 bar + readability of UUT
DC Voltage: ± 120 ppm + readability of UUT
DC Current: $\pm 0.08\%$ + readability of UUT

Comments: The reported results fall within the manufacturers specification

The SI unit of pressure is the pascal (Pa) ... 1MPa=10bar=145.038lbf/in²=10.1972kgf/cm². 1kPa=10mbar=0.2953inHg=7.50062mmHg=4.01463inH₂O.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to the units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, without the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

UKAS ACCREDITED LABORATORY No: 0495

CERTIFICATE No:

311689

Page 2 of 2

Range Max:
Range Min:

Resolution:
Readability:

Units:

Applied Pressure [1]	UUT Reading [2]	Deviation [2] - [1]	Deviation as % FSD
0.05000	0.0510	0.0010	0.014
0.50000	0.5008	0.0008	0.011
1.00000	1.0008	0.0008	0.011
2.00991	2.0108	0.0009	0.013
3.01041	3.0113	0.0009	0.013
4.01083	4.0116	0.0008	0.011
5.01130	5.0120	0.0007	0.010
6.01179	6.0124	0.0006	0.009
7.01223	7.0127	0.0005	0.007
6.01182	6.0125	0.0007	0.010
5.01127	5.0121	0.0008	0.012
4.01090	4.0118	0.0009	0.013
3.01046	3.0114	0.0009	0.013
2.00992	2.0109	0.0010	0.014
1.00000	1.0009	0.0009	0.013
0.50000	0.5009	0.0009	0.013
0.05000	0.0510	0.0010	0.014

DC Voltage Accuracy

Before Adjustment UUT Reading (Vdc) [1]	Actual Applied Voltage (Vdc) [2]	Deviation [2] - [1]
40.000	40.0060	0.0060
25.000	25.0040	0.0040
10.000	10.0010	0.0010
-10.000	-10.0020	-0.0020
-30.000	-30.0050	-0.0050

DC Current Accuracy

Before Adjustment UUT Reading (mAdc) [1]	Actual Applied Current (mAdc) [2]	Deviation [2] - [1]
10.000	9.9990	-0.0010
25.000	24.9970	-0.0030
50.000	50.0020	0.0020